


# Atrazine Plate, Detailed ELISA Procedure


## 1. Addition of Assay Buffer

Add 25  $\mu$ l of the Assay Buffer into the wells of the test strips according to the working scheme given. We recommend using duplicates or triplicates.




## 2. Addition of Standards, Samples

Add 25  $\mu$ l of the standard solutions, control or samples into the wells of the test strips according to the working scheme given. We recommend using duplicates or triplicates.




## 3. Addition of Enzyme Conjugate

Add 50  $\mu$ l of the enzyme conjugate to the individual wells successively using a multi-channel pipette or a stepping pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop for 30 seconds. Be careful not to spill contents. Incubate for 30 minutes at room temperature.




## 4. Washing of Plates

After incubation, remove the covering and vigorously shake the contents of the wells into a sink. Wash the strips three times with a multi-channel pipette or wash bottle using the 1X washing buffer solution. Please use at least a volume of 250  $\mu$ l of washing buffer for each well and each washing step. Remaining buffer in the wells should be removed by patting the plate dry on a stack of paper towels.




## 5. Addition of Substrate/Color Solution

Add 100  $\mu$ l of substrate/color solution to the individual wells successively using a multi-channel pipette or a stepping pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop. Be careful not to spill contents. Incubate the strips for 15-20 min at room temperature.




## 6. Addition of Stopping Solution

Add 50  $\mu$ l of stop solution to the wells in the same sequence as for the substrate solution using a multi-channel pipette or a stepping pipette.



## 7. Measurement of Color

Read the absorbance at 450 nm using a microplate ELISA reader. Calculate results.




For Ordering or Technical Assistance Contact:  
ABRAXIS, LLC  
54 Steamwhistle Drive, Warminster, PA 18974  
Phone: 215-357-3911 Fax: 215-357-5232  
[www.abraxiskits.com](http://www.abraxiskits.com)

Atrazine Plate Kit Part # 520005

# Atrazine Plate, Concise ELISA Procedure


## 1. Addition of Assay Buffer

Add 25  $\mu$ L of the Assay Buffer.




## 2. Addition of Standards, Samples

Add 25  $\mu$ L of standard solutions, control or samples.




## 3. Addition of Enzyme Conjugate

Add 50  $\mu$ L of enzyme conjugate. Cover and mix for 30 seconds by rotating on benchtop. Incubate for 30 minutes at room temperature.




## 4. Washing of Plates

Wash the plates three times with 250  $\mu$ L of 1X washing buffer.




## 5. Addition of Substrate/Color Solution

Add 100  $\mu$ L of substrate/color solution. Incubate 15-20 minutes at room temperature and away from direct sunlight.




## 6. Addition of Stopping Solution

Add 50  $\mu$ L of stop solution.



## 7. Measurement of Color

Measure color at 450 nm.  
Calculate results.



For Ordering or Technical Assistance Contact:  
ABRAXIS, LLC  
54 Steamwhistle Drive, Warminster, PA 18974  
Phone: 215-357-3911 Fax: 215-357-5232  
[www.abraxiskits.com](http://www.abraxiskits.com)

Atrazine Plate Kit Part # 520005