


Gentamicin Plate, Detailed ELISA Procedure


1. Addition of Assay Buffer

Add 25 μ L of Assay Buffer to the wells of the test strips successively using a multi-channel pipette or a stepping pipette according to the working scheme given. We recommend using duplicates or triplicates.




2. Addition of Standards, Samples

Add 25 μ L of the standard solutions and samples to the wells of the test strips according to the working scheme given.




3. Addition of Conjugate Solution

Add 100 μ L of the enzyme conjugate to the individual wells successively using a multi-channel pipette or a stepping pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop. Be careful not to spill contents. Incubate the strips for 30 min at room temperature.




4. Washing of Plates

After incubation, remove the covering and vigorously shake the contents of the wells into a sink. Wash the strips three times with a multi-channel pipette or wash bottle using the diluted 1X washing buffer solution. Please use at least a volume of 250 μ L of washing buffer for each well and each washing step. Remaining buffer in the wells should be removed by patting the plate dry on a stack of paper towels.




5. Addition of Substrate/Color Solution

Add 100 μ L of substrate/color solution to the individual wells successively using a multi-channel pipette or a stepping pipette. Cover the wells with parafilm or tape and mix the contents by moving the strip holder in a rapid circular motion on the benchtop. Be careful not to spill contents. Incubate the strips for 15 min at room temperature.




6. Addition of Stopping Solution

Add 100 μ L of stop solution to the wells in the same sequence as for the substrate solution using a multi-channel pipette or a stepping pipette.



7. Measurement of Color

Read the absorbance at 450 nm using a microplate ELISA reader. Calculate results.




For Ordering or Technical Assistance Contact:
ABRAXIS, LLC
54 Steamwhistle Drive, Warminster, PA 18974
Phone: 215-357-3911 Fax: 215-357-5232
www.abraxiskits.com

Gentamicin Plate, Concise ELISA Procedure


1. Addition of Assay Buffer

Add 25 μ L of Assay Buffer.




5. Addition of Substrate/Color Solution

Add 100 μ L of substrate/color solution. Incubate 15 minutes at room temperature and away from direct sunlight.




2. Addition of Standards, Samples

Add 25 μ L of the standard solutions and samples.




6. Addition of Stopping Solution

Add 100 μ L of stop solution.




3. Addition of Conjugate Solution

Add 100 μ L of the enzyme conjugate. Cover and mix for 30 seconds by rotating on benchtop. Incubate for 30 min at room temperature.





7. Measurement of Color

Measure color at 450 nm. Calculate results.



4. Washing of Plates

Wash the plates three times with 250 μ L of diluted 1X washing buffer.



For Ordering or Technical Assistance Contact:
ABRAKIS, LLC
54 Steamwhistle Drive, Warminster, PA 18974
Phone: 215-357-3911 Fax: 215-357-5232
www.abraxiskits.com